## **GSM28 Medical** 28 Watt Global Performance Switchers



### SPECIFICATIONS:

#### Ac Input

85-264 Vac, 47-63 Hz single phase.

#### Input Current

Maximum input current at 120 Vac, 60 Hz with full rated output load is 0.85 A.

#### Hold-up Time

15 ms minimum from loss of ac input at full load, nominal line (120 Vac).

#### **Output Power**

Normal continuous output power is 28 W, 32 W peak for 60 s maximum duration, 10% duty cycle. Factory set to begin power limiting at approximately 35 W.

#### **Overload Protection**

Fully protected against short circuit and output overload. Short circuit protection is cycling type power limit.

#### **Output Noise**

0.5% rms, 1% pk-pk, 20 MHz bandwidth, differential mode. Measured with scope probe directly across output terminals of the power supply with load terminated with 0.1 uF capacitor.

#### **Overvoltage Protection**

Built in with firing point set per ratings table. OVP firing reduces voltage to less than 50% of nominal voltage in 50 ms.

#### Voltage Adjustment

Factory set with fixed resistors to maximize reliability.

#### Efficiency

70% minimum for the 5.1 V model at full rated load, nominal input voltage. Efficiency increases as output voltage increases.

#### **Input Protection**

Internal ac fuse provided on all units. Designed to open only if a catastrophic failure occurs in the unit.

#### Inrush Current

Inrush limited by internal thermistor. The inrush at 230 Vac, averaged over the first ac half-cycle under cold start conditions will not exceed 32 A.

## FEATURES:

- · Industry's smallest 28 W medically approved switcher
- Compact size (4.00" x 2.59" x 0.92")
- Wide-range ac input: 85-264 Vac
- Less than 25 µA leakage current @ 120 Vac
- Approved to UL2601-1, EN60601-1
- EMI to FCC, CISPR 11 Class B/IEC601-1-2
- Overvoltage protection standard
- **(** marked to LVD

#### **Temperature Coefficient**

0.03%/°C typical on all outputs.

#### Environmental

Designed for 0 to 50°C operation at full rated output power; derate output current and total output power by 2.5% per °C above 50°C.

#### Medical EMI/EMC Compliance

All models include built-in EMI filtering to meet the following emissions requirements:

COMPLIANCE LEVEL
EN55011 Class B; FCC Class B
EN61000-4-2, 6 kV contact, 8 kV air
EN61000-4-3, 3 V/meter
EN61000-4-4, 2 kV, 5 kHz
EN61000-4-5, 1 kV diff., 2 kV com.

#### Earth Leakage Current

Leakage current measured in the Gnd wire connection when measured per EN60601-1 or UL2601-1 is as follows:

Medical Model	Normal Leakage	Single Fault Leakage	Test Voltage	Test Method
GSM28	25 A	45 A	132 Vac/60 Hz	UL2601-1
GSM28	50 A	90 A	264 Vac/50 Hz	IEC60601-1

#### **Medical Medical Safety**

Condor D.C. Power Supplies, Inc. declares under our sole responsibility that all GSM models are in conformity with the applicable requirements of UL2601-1 Patient Care Equipment, CSA-C22.2 No. 234 (with additional tests to C22.2 No. 601.1 per T.I.L. CA-08), EN60601-1.

# **GSM28 Medical 28 Watt Single Output**

Medical Model	Output	Current	Load Regulation	Initial Setpoint Tolerance	OVP Setpoint	Ripple and Noise
GSM28-5	5.1 V	5.5 A	0.75%	2.5%	$6.2 \pm 0.6 \text{ V}$	1.4%
GSM28-12	12 V	2.3 A	0.75%	2.5%	14 ± 1.0 V	1%
GSM28-15	15 V	1.9 A	0.75%	2.5%	18.5 ± 1.5 V	1%
GSM28-24	24 V	1.2 A	0.75%	2.5%	28 ± 2.5 V	1%
GSM28-28	28 V	1.0 A	0.75%	2.5%	34 ± 2.8 V	1%

## **GSM28** MECHANICAL SPECIFICATIONS



Environmental Specification	Operating	Non-operating
Temperature (A)	See individual specs	-40 to +85°C
Humidity (A)	0 to 95% RH	0 to 95% RH
Shock (B)	20 g <sub>pk</sub>	40 g <sub>pk</sub>
Altitude	-500 to 10,000 ft	-500 to 40,000 ft
Vibration (C)	1.5 g <sub>rms</sub> , 0.003 g²/Hz	5 g <sub>rms</sub> , 0.026 g²/Hz

A. Units should be allowed to warm up/operate under non-condensing conditions before application of power.

Random vibration-10 to 2000Hz, 6dB/octave roll-off from 350 to 2000Hz, 3 В.

orthogonal axes. Tested for 10 min./axis operating and 1 hr./axis non-operating. C. Shock testing—half-sinusoidal, 10  $\pm$  3 ms duration,  $\pm$  direction, 3 orthogonal axes, total 6 shocks.

